

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00815	Calib. Date:	09/05/2019
Certification File No.:	00816	Cert. Date:	09/05/2019
Linearity File No.:	00817	Lin. Date:	09/05/2019
Solution File No.:	00844	Soln. Date:	01/10/2020
Sequential File No.:	00844	File Date:	01/10/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	1125

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:24S		
Control 1 EC	0.100%	21:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:25S		
Control 2 EC	0.100%	21:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27S		
Control 3 EC	0.100%	21:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PETERSON

First Name: PTLM BRIAN

MI: D

Signature: _____

[Handwritten Signature] #129

Badge No.: 129

Date: 01/10/2020

