

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00706	Calib. Date: 01/09/2018
Certification File No.:	00707	Calib. No.: 00037
Linearity File No.:	00708	Cert. Date: 01/09/2018
Solution File No.:	00724	Cert. No.: 00028
Sequential File No.:	00724	Lin. Date: 01/09/2018
		Lin. No.: 00032
		Soln. Date: 04/30/2018
		Soln. No.: 00179
		File Date: 04/30/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0089

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:18D		
Control 1 EC	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		
Control 2 EC	0.099%	09:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		
Control 3 EC	0.099%	09:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: _____



Badge No.: 090

Date: 04/30/2018

