

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00633 Calib. Date: 09/22/2016 Calib. No.: 00033
Certification File No.: 00634 Cert. Date: 09/22/2016 Cert. No.: 00024
Linearity File No.: 00635 Lin. Date: 09/22/2016 Lin. No.: 00028
Solution File No.: 00653 Soln. Date: 01/04/2017 Soln. No.: 00151
Sequential File No.: 00653 File Date: 01/04/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 05/17/2018
Solution Control Lot: 201605D Bottle No.: 0202

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:21S		
Control 1 EC	0.100%	07:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23S		
Control 2 EC	0.098%	07:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 3 EC	0.099%	07:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ARMINIO

First Name: GERALD

MI: A

Signature: _____

Badge No.: 80

Date: 01/04/2017

JAN 05 2017