

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00815	Calib. Date:	09/05/2019
Certification File No.:	00816	Cert. Date:	09/05/2019
Linearity File No.:	00817	Lin. Date:	09/05/2019
Solution File No.:	00826	Soln. Date:	10/17/2019
Sequential File No.:	00826	File Date:	10/17/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0139

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:25D		
Control 1 EC	0.098%	11:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 2 EC	0.099%	11:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 3 EC	0.099%	11:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CHAPLINSKI

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_



Badge No.: 061

Date: 10/17/2019

