

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00582	Calib. Date: 10/16/2015
Certification File No.:	00583	Calib. No.: 00030
Linearity File No.:	00584	Cert. Date: 10/16/2015
Solution File No.:	00601	Cert. No.: 00021
Sequential File No.:	00601	Lin. Date: 10/16/2015
		Lin. No.: 00025
		Soln. Date: 03/09/2016
		Soln. No.: 00137
		File Date: 03/09/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 0363

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:59S		
Control 1 EC	0.101%	21:00S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:00S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:01S		
Control 2 EC	0.101%	21:01S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:01S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02S		
Control 3 EC	0.100%	21:03S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:03S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GLENNY

First Name: DAVID

MI: -

Signature: \_\_\_\_\_



Badge No.: 71

Date: 03/09/2016

