

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00657	Calib. Date: 02/02/2017
Certification File No.:	00658	Cert. No.: 00034
Linearity File No.:	00659	Cert. Date: 02/02/2017
Solution File No.:	00669	Lin. No.: 00025
Sequential File No.:	00669	Lin. Date: 02/02/2017
		Lin. No.: 00029
		Soln. Date: 05/12/2017
		Soln. No.: 00159
		File Date: 05/12/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%			Expires:	05/17/2018
Solution Control Lot:	201605D			Bottle No.:	0206

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:40D		
Control 1 EC	0.103%	14:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41D		
Control 2 EC	0.102%	14:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	14:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43D		
Control 3 EC	0.102%	14:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	14:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____



Badge No.: 96

Date: 05/12/2017

