

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00788	Calib. Date:	04/15/2019
Certification File No.:	00789	Cert. Date:	04/15/2019
Linearity File No.:	00790	Lin. Date:	04/15/2019
Solution File No.:	00811	Soln. Date:	08/12/2019
Sequential File No.:	00811	File Date:	08/12/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0607

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:47D		
Control 1 EC	0.097%	13:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		
Control 2 EC	0.097%	13:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	13:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		
Control 3 EC	0.097%	13:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STATTEL

First Name: DAVID


MI: K

Signature: _____



Badge No.: 116

Date: 08/12/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00788	Calib. Date:	04/15/2019
Certification File No.:	00789	Cert. Date:	04/15/2019
Linearity File No.:	00790	Lin. Date:	04/15/2019
Solution File No.:	00810	Soln. Date:	08/12/2019
Sequential File No.:	00810	File Date:	08/12/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0488

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:07D		
Control 1 EC	-.---%	12:07D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	12:07D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STATTEL


First Name: DAVID

MI: K

Signature: 

Badge No.: 116

Date: 08/12/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00788	Calib. Date:	04/15/2019
Certification File No.:	00789	Cert. Date:	04/15/2019
Linearity File No.:	00790	Lin. Date:	04/15/2019
Solution File No.:	00809	Soln. Date:	08/12/2019
Sequential File No.:	00809	File Date:	08/12/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0187

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:53D		
Control 1 EC	-.---%	09:54D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	09:54D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STATTEL

First Name: DAVID

MI: K

Signature: *Atm* 

Badge No.: 116

Date: 08/12/2019

