

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00413 Calib. Date: 04/04/2013 Calib. No.: 00019
Certification File No.: 00414 Cert. Date: 04/04/2013 Cert. No.: 00012
Linearity File No.: 00415 Lin. Date: 04/04/2013 Lin. No.: 00015
Solution File No.: 00426 Soln. Date: 07/02/2013 Soln. No.: 00092
Sequential File No.: 00426 File Date: 07/02/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0388

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 12:18D | | |
| Control 1 EC | 0.098% | 12:19D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 12:19D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:20D | | |
| Control 2 EC | 0.097% | 12:20D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 12:20D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:21D | | |
| Control 3 EC | 0.096% | 12:22D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 12:22D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:22D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LAYMAN

First Name: JOSHUA

MI: E

Signature: _____



Badge No.: 101

Date: 07/02/2013

