

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00927 Calib. Date: 03/09/2021 Calib. No.: 00046
Certification File No.: 00928 Cert. Date: 03/09/2021 Cert. No.: 00036
Linearity File No.: 00929 Lin. Date: 03/09/2021 Lin. No.: 00040
Solution File No.: 00932 Soln. Date: 03/30/2021 Soln. No.: 00247
Sequential File No.: 00932 File Date: 03/30/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 1377

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:09S		
Control 1 EC	0.098%	15:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:10S		
Control 2 EC	0.099%	15:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12S		
Control 3 EC	0.099%	15:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature:  #129

Badge No.: 129

Date: 03/30/2021

