

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00927 Calib. Date: 03/09/2021 Calib. No.: 00046
Certification File No.: 00928 Cert. Date: 03/09/2021 Cert. No.: 00036
Linearity File No.: 00929 Lin. Date: 03/09/2021 Lin. No.: 00040
Solution File No.: 00936 Soln. Date: 04/12/2021 Soln. No.: 00248
Sequential File No.: 00936 File Date: 04/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 1365

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:23D		
Control 1 EC	0.090%	10:23D	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.091%	10:23D	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:24D		
Control 2 EC	0.091%	10:25D	33.8°C	CONTROL TEST FAILED
Control 2 IR	0.092%	10:25D	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:26D		
Control 3 EC	0.092%	10:26D	33.8°C	CONTROL TEST FAILED
Control 3 IR	0.093%	10:26D	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:27D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: _____

Kenneth M. Marrone #69

Badge No.: 069

Date: 04/12/2021

