

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 01000 Calib. Date: 06/01/2022 Calib. No.: 00049
Certification File No.: 01001 Cert. Date: 06/01/2022 Cert. No.: 00039
Linearity File No.: 01002 Lin. Date: 06/01/2022 Lin. No.: 00043
Solution File No.: 01010 Soln. Date: 06/17/2022 Soln. No.: 00272
Sequential File No.: 01010 File Date: 06/17/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0811

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:35D		
Control 1 EC	0.101%	19:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:36D		
Control 2 EC	0.100%	19:37D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:38D		
Control 3 EC	0.100%	19:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROUNDS

First Name: TAYLOR

MI: W

Signature: _____



Badge No.: 133

Date: 06/17/2022

