

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00788 Calib. Date: 04/15/2019 Calib. No.: 00041
Certification File No.: 00789 Cert. Date: 04/15/2019 Cert. No.: 00031
Linearity File No.: 00790 Lin. Date: 04/15/2019 Lin. No.: 00035
Solution File No.: 00801 Soln. Date: 07/02/2019 Soln. No.: 00202
Sequential File No.: 00801 File Date: 07/02/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0462

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:37D | | |
| Control 1 EC | 0.098% | 07:37D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 07:37D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:38D | | |
| Control 2 EC | 0.098% | 07:39D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 07:39D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:39D | | |
| Control 3 EC | 0.098% | 07:40D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 07:40D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:40D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GARGUILLO

First Name: JAVIER

MI: A

Signature: _____

Badge No.: 84

Date: 07/02/2019

