

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00739	Calib. Date: 07/02/2018
Certification File No.:	00740	Calib. No.: 00038
Linearity File No.:	00741	Cert. Date: 07/02/2018
Solution File No.:	00758	Cert. No.: 00029
Sequential File No.:	00758	Lin. Date: 07/02/2018
		Lin. No.: 00033
		Soln. Date: 09/07/2018
		Soln. No.: 00186
		File Date: 09/07/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:14D		
Control 1 EC	0.100%	09:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16D		
Control 2 EC	0.099%	09:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:17D		
Control 3 EC	0.099%	09:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 090

Date: 09/07/2018

