

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00681	Calib. Date:	07/27/2017
Certification File No.:	00682	Cert. Date:	07/27/2017
Linearity File No.:	00683	Lin. Date:	07/27/2017
Solution File No.:	00693	Soln. Date:	10/19/2017
Sequential File No.:	00693	File Date:	10/19/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	0896

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:47D		
Control 1 EC	0.100%	08:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48D		
Control 2 EC	0.100%	08:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49D		
Control 3 EC	0.100%	08:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: *Cpt Kenneth #69*

Badge No.: 069

Date: 10/19/2017

