

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00582 Calib. Date: 10/16/2015 Calib. No.: 00030  
Certification File No.: 00583 Cert. Date: 10/16/2015 Cert. No.: 00021  
Linearity File No.: 00584 Lin. Date: 10/16/2015 Lin. No.: 00025  
Solution File No.: 00590 Soln. Date: 11/15/2015 Soln. No.: 00133  
Sequential File No.: 00590 File Date: 11/15/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 04/13/2017  
Solution Control Lot: 15D138 Bottle No.: 0361

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:04S		
Control 1 EC	0.102%	23:04S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:04S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05S		
Control 2 EC	0.101%	23:06S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:06S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06S		
Control 3 EC	0.101%	23:07S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:07S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GLENNY

First Name: DAVID

MI: -

Signature: \_\_\_\_\_

*David Glenny #71*

Badge No.: 71

Date: 11/15/2015

NOV 24 2015