

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE	
Calibration File No.: 00681	Calib. Date: 07/27/2017
Certification File No.: 00682	Calib. No.: 00036
Linearity File No.: 00683	Cert. Date: 07/27/2017
Solution File No.: 00700	Cert. No.: 00027
Sequential File No.: 00700	Lin. Date: 07/27/2017
	Lin. No.: 00031
	Soln. Date: 11/27/2017
	Soln. No.: 00172
	File Date: 11/27/2017

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXC S3-0164
Control Solution %: 0.100%		Expires: 09/14/2019
Solution Control Lot: 17290		Bottle No.: 0307

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:04S		
Control 1 EC	0.102%	07:05S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:05S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:05S		
Control 2 EC	0.100%	07:06S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:06S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:07S		
Control 3 EC	0.101%	07:07S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:07S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: *Kenneth Marrone* #069

Badge No.: 069

Date: 11/27/2017