

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00890	Calib. Date:	09/18/2020
Certification File No.:	00891	Cert. Date:	09/18/2020
Linearity File No.:	00892	Lin. Date:	09/18/2020
Solution File No.:	00909	Soln. Date:	12/02/2020
Sequential File No.:	00909	File Date:	12/02/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	1364

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:22S		
Control 1 EC	0.098%	18:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:24S		
Control 2 EC	0.098%	18:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:25S		
Control 3 EC	0.099%	18:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STATTEL

First Name: DAVID

MI: K

Signature: _____



Badge No.: 116

Date: 12/02/2020

