

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00633 Calib. Date: 09/22/2016 Calib. No.: 00033  
Certification File No.: 00634 Cert. Date: 09/22/2016 Cert. No.: 00024  
Linearity File No.: 00635 Lin. Date: 09/22/2016 Lin. No.: 00028  
Solution File No.: 00648 Soln. Date: 12/21/2016 Soln. No.: 00149  
Sequential File No.: 00648 File Date: 12/21/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 04/13/2017  
Solution Control Lot: 15D138 Bottle No.: 0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:33S		
Control 1 EC	0.099%	23:34S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:34S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34S		
Control 2 EC	0.097%	23:35S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:35S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:36S		
Control 3 EC	0.098%	23:36S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:36S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ARMINIO

First Name: GERALD

MI: A

Signature: \_\_\_\_\_

Badge No.: 080

Date: 12/21/2016

DEC 28 2016