

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARXA-0069 |
| Location: | MARLBORO TWP. POLICE | | |
| Calibration File No.: | 00512 | Calib. Date: | 10/09/2014 |
| Certification File No.: | 00513 | Cert. Date: | 10/09/2014 |
| Linearity File No.: | 00515 | Lin. Date: | 10/09/2014 |
| Solution File No.: | 00516 | Soln. Date: | 10/09/2014 |
| Sequential File No.: | 00516 | File Date: | 10/09/2014 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDXC S3-0162 |
| Solution Control Lot: | 13K125 | Expires: | 11/12/2015 |
| | | Bottle No.: | 1233 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 17:43D | | |
| Control 1 EC | 0.100% | 17:43D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 17:43D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:44D | | |
| Control 2 EC | 0.100% | 17:45D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 17:45D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:45D | | |
| Control 3 EC | 0.100% | 17:46D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 17:46D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:46D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-066 (4)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: *[Handwritten Signature]*

Badge No.: 5925

Date: 10/09/2014

