

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00512	Calib. Date: 10/09/2014
Certification File No.:	00513	Calib. No.: 00024
Linearity File No.:	00515	Cert. Date: 10/09/2014
Solution File No.:	00532	Cert. No.: 00016
Sequential File No.:	00532	Lin. Date: 10/09/2014
		Lin. No.: 00020
		Soln. Date: 01/17/2015
		Soln. No.: 00119
		File Date: 01/17/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1246

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:51S		
Control 1 EC	0.102%	21:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:52S		
Control 2 EC	0.101%	21:53S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:53S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:54S		
Control 3 EC	0.100%	21:54S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:54S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ARMINIO

First Name: GERALD

MI: A

Signature: \_\_\_\_\_



Badge No.: 080

Date: 01/17/2015

